

Barcelona, 25<sup>th</sup> of November of 2008

To whom it may concern

**SUBJECT: BIOCOMPATIBILITY TESTS FOR ANDROPENIS GOLD**

The following tables summarize the results of tests conducted on behalf of ANDROMEDICAL S.A. on the test items ANDROPENIS, ANDROPENIS GOLD and ANDRORING by Biolab S.p.A. in the Assay Center of Via B.Buozzi, 2 – 20090 Vimodrone Milan (Italy) from the 2<sup>nd</sup> of September of 2005 to the 24<sup>th</sup> of November of 2008.

Details of the analytical methods, samples analyzed and sample preparation, test conditions and procedures are described in the final reports referenced in each table. A copy of those reports is retained in the archives of the Assay Center.

**Test Item: ANDROPENIS GOLD**

HEAVY METALS DETERMINATION (EN 71-3:1994/AC:2002) (see report 2008-910-2iSAM)

metals	limit EN 71- 3:1994/AC:200 2 (mg/kg)	metal bar (mg/kg)	axis (mg/kg)	superior plastic support (mg/kg)	silicone band (mg/kg)	andro top (mg/kg)
Barium	1000	<100	<100	<100	<100	<100
Cadmium	75	<7.5	<7.5	<7.5	<7.5	<7.5
Chromium	60	<6	<6	<6	<6	<6
Lead	90	10	<9	<9	<9	<9
Mercury	60	<6	<6	<6	<6	<6
Arsenic	25	<2.5	<2.5	<2.5	<2.5	<2.5
Antimony	60	<6	<6	<6	<6	<6
Selenium	500	<50	<50	<50	<50	<50
<i>Copper</i>	-	<2.5	<2.5	<2.5	<2.5	<2.5
<i>Nickel</i>	-	96	<2.5	<2.5	<2.5	<2.5
<i>Zinc</i>	-	<2.5	<2.5	<2.5	<2.5	<2.5
<i>Gold</i>	-	<2.5	<2.5	<2.5	<2.5	<2.5

**BIOLOGICAL TESTS (ISO 10993)**

<b>Model</b>	<b>Test</b>	<b>Sample</b>	<b>Result</b>	<b>Report</b>
ANDROPENIS	Cytotoxicity – Elution Test (ISO 10993 – 5: 1999)	PLASTIC	NOT CYTOTOXIC	SAM2817i-1
		FOAM	NOT CYTOTOXIC	
		ALLUMINUM ALLOY	NOT CYTOTOXIC	
		SILICONE	NOT CYTOTOXIC	
	Delayed Type Hypersensitivity Test (ISO 10993- 10 : 2002)	POOL OF THE COMPONENTS. (PLASTIC, FOAM, ALLUMINUM	NOT SENSITIZING	SAM2817i-2
Skin Irritation Test (ISO 10993-10: 2002)	ALLOY, SILICONE, METAL BAR)	NOT IRRITANT	SAM2817i-3	
ANDROPENIS GOLD	Cytotoxicity – Elution Test (ISO 10993 – 5: 1999)	METAL BAR	NOT CYTOTOXIC	2008-910-2i SAM

**Test Item: ANDRORING**
**BIOLOGICAL TESTS (ISO 10993)**

<b>Model</b>	<b>Test</b>	<b>Sample</b>	<b>Result</b>	<b>Report</b>
ANDRORING	Cytotoxicity – Elution Test (ISO 10993 – 5: 1999)	ANDRORING	NOT CYTOTOXIC	SAM3036i-2

Biolab Española SL  
Luca Salvi  
Managing Director